

<b>Comparison of Lead Apron Protection</b>			
<b>Distance (ft)</b>	<b>Thickness (mm)</b>	<b>Angle (deg)</b>	<b>Dose (mR)</b>
<b>2</b>	<b>0.5</b>	<b>90</b>	<b>0.012</b>
	<b>0.25</b>	<b>90</b>	<b>1.595</b>
<b>3</b>	<b>0.5</b>	<b>90</b>	<b>0</b>
	<b>0.25</b>	<b>90</b>	<b>0.834</b>
<b>4</b>	<b>0.5</b>	<b>90</b>	<b>0</b>
	<b>0.25</b>	<b>90</b>	<b>0.546</b>
<b>5</b>	<b>0.5</b>	<b>90</b>	<b>0</b>
	<b>0.25</b>	<b>90</b>	<b>0.338</b>
<b>6</b>	<b>0.5</b>	<b>90</b>	<b>0</b>
	<b>0.25</b>	<b>90</b>	<b>0</b>
<b>2</b>	<b>0.5</b>	<b>60</b>	<b>0</b>
	<b>0.25</b>	<b>60</b>	<b>1.057</b>
<b>3</b>	<b>0.5</b>	<b>60</b>	<b>0</b>
	<b>0.25</b>	<b>60</b>	<b>0.62</b>
<b>4</b>	<b>0.5</b>	<b>60</b>	<b>0</b>
	<b>0.25</b>	<b>60</b>	<b>0.389</b>
<b>5</b>	<b>0.5</b>	<b>60</b>	<b>0</b>
	<b>0.25</b>	<b>60</b>	<b>0</b>

130 X's more radiation

The cumulative dose through a lead apron during 1 minute of fluoroscopy at various distances. A phantom is supine with 6 in of polyethylene blocks under it. In room fluoro II is 4in above midline of phantom. The ion chamber records measurements perpendicular (90 deg) to the migsagittal plane of the phantom and 60 degrees off of perpendicular.



Digital Radiography Solutions "Specializing in Educational Conferences and Seminars"

**Disclaimer** - Digital Radiography Solutions (DRS) cannot be held responsible for any images taken or passed in your facility. Your radiologist always has the final say if an image is diagnostic.