

CESIUM DR TECHNIQUE CHART 33% More mAs (500 Speed)

Part	View	Small		Medium		Large	
		kV	mAs	kV	mAs	kV	mAs
Knee	AP/Lat (grid)	85	3.3	85	4	85	4.7
Knee	AP/Lat (non grid)	70	2.7	70	3.3	70	4
Knee	Sunrise (non grid)	70	3.3	70	4	70	4.7
L-Spine	AP (grid)	90	9.3	90	13.3	90	19
L-Spine	Oblique (grid)	90	15	90	21	90	29
L-Spine	X-Table Lat (grid)	95	21	95	29	95	40
Mandible	PA (grid)	85	4	85	5.3	85	6.7
Mandible	Obl (grid - 40")	81	3.1	81	4.3	81	5.5
Mandible	Obl (non grid - 40")	70	2.7	70	3.7	70	4.8
Nasal Bones	Lateral (non grid)	70	2	70	2.5	70	3.1
Pelvis	AP (grid)	85	8	85	12	85	16
Ribs	PA Upper (72")	85	8	85	10.7	85	13.3
Ribs	Obl Upper (72")	85	12	85	16	85	20
Ribs	AP Lower (40")	85	8	85	12	85	16
Shoulder	AP (grid)	85	4.1	85	6	85	8
Shoulder	AP (non grid)	70	2.7	70	3.6	70	4.4
Shoulder	Y View (grid)	85	11	85	15	85	21
Shoulder	Axillary (non grid)	70	2.7	70	3.7	70	4.7
Sinus	Caldwell	85	4.7	85	5.9	85	7.1
Sinus	Waters	85	5.3	85	6.7	85	8
Sinus	Lateral	85	2.4	85	3.1	85	3.9
Skull	PA	85	4.4	85	5.6	85	6.8
Skull	Lat (grid)	85	3.3	85	4	85	4.7
Tib-Fib	AP/Lat (non grid)	70	1.9	70	2.5	70	3.1
Toe	All Views	66	0.7	66	0.8	66	0.9
T-Spine	AP	85	7.3	85	10	85	12.7
T-Spine	Lat (breathing)	90	17	90	24	90	31
Wrist	PA	66	0.8	66	1.1	66	1.3
Wrist	Lat	70	1.1	70	1.3	70	1.6
Zygomatic Arch	AP Axial - Townes	80	3.5	80	4.3	80	4.8
Zygomatic Arch	SMV (grid)	80	3.3	80	4.1	80	4.9
Zygomatic Arch	SMV (non grid)	70	2.8	70	3.6	70	4.4



Digital Radiography Solutions "Maximizing Image Quality, Minimizing Patient Dose"

Disclaimer: Digital Radiography Solutions (dRs) cannot be held responsible for any images taken or passed at your facility. Your radiologist or physician always has the final say if the image is of diagnostic value.